

INVITATION TO BID

State of Ohio
Department of Administrative Services
General Services Division
Office of Procurement Services

The Original Signed Bid must be submitted to the Office of Procurement Services to receive consideration for award.		BIDDER NAME	
BID NUMBER OT904316	OPENING DATE (1:00 p.m.) February 22, 2016	STREET ADDRESS <input type="checkbox"/> Check if remit address is different and list on separate sheet	
General Services Division Office of Procurement Services 4200 Surface Road Columbus, OH 43228-1395 Attn: Bid Desk		CITY STATE ZIP	
		COUNTY	MBE/EDGE CERTIFICATE NUMBER
		TELEPHONE NO. ()	TOLL FREE NO. 1 - ()
		CONTACT PERSON	FAX NO. ()
REQ./INDEX NO. DRC012	BID NOTICE DATE 01/13/16	CONTRACTOR'S E-MAIL ADDRESS	
SELECT YOUR PREFERRED METHOD OF RECEIVING PURCHASE ORDERS AND ENTER THE E-MAIL OR FAX NUMBER INFORMATION (ONLY SELECT ONE METHOD) <input type="checkbox"/> E-Mail <input type="checkbox"/> Fax			
In addition to the standard terms for payment, the payment terms for state agency(ies) will be 2%, 10 Days, Net 30 Days unless otherwise stated in the following space. If no discount is offered, bidder should circle "Net 30 Days". ____%, ____Days, Net 30 Days			
<u>PARTICIPATING AGENCY(IES):</u> OHIO DEPARTMENT OF REHABILITATION AND CORRECTION, 770 WEST BROAD STREET, COLUMBUS, OH 43222			
ADDENDUM FOR CHANGE ADDENDUM NO.: 3 REVISION DATE: 02/08/16			
<u>THE DEPARTMENT OF ADMINISTRATIVE SERVICES, OFFICE OF PROCUREMENT SERVICES IS SOLICITING BIDS FOR:</u> DRUG TESTING CUPS AND ORAL FLUID TESTERS FOR THE OHIO DEPARTMENT OF REHABILITATION AND CORRECTION			
Attached is page 5 to this Invitation to Bid. Remove the corresponding page from the existing bid and replace with the attached.			
As indicated herein this addendum is issued to change the chemical name for OXY to: Oxycodone on page 5, II, A, 1, i.			

SPECIFICATIONS

I. SCOPE

The State of Ohio, Department of Administrative Services (DAS) on behalf of the Ohio Department of Rehabilitation and Correction (DRC) and a division of the DRC, the Ohio Adult Parole (APA), is soliciting bids for a Contractor to provide drug testing cups and oral fluid testers for delivery to DRC and APA facilities throughout Ohio. There are more than twenty (20) State facilities throughout Ohio that will order and receive these items.

The Contract period is for March 1, 2016 through February 28, 2018, unless renewed or terminated.

II. SPECIFICATIONS.

A. ITEM 1 – 8-PANEL DRUG TESTING CUP

1. The drug testing cups will meet the following specifications:
 - a. The test strips must be single drug per test.
 - b. The drug testing cup must be an all-inclusive cup and only require the amount of urine needed for the collection and accurate results. No other process or device can be required to perform the test.
 - c. The drug testing cup will have easy readable window of test strips and results.
 - d. The drug testing cup will have a temperature strip to ensure the sample is within normal temperature ranges.
 - e. The drug testing cup results will be available to read within five (5) minutes of collection.
 - f. The drug test containers should not have a max fill-line on the containers.
 - g. All products shipped must have a minimum twelve (12) month shelf-life after receipt by the State and each individual package must be dated.
 - h. Drug test cups shall seal in such a way as to prevent any leakage during shipment to the testing facility.
 - i. The following items are the drugs to be tested and each of their screening levels:
 - AMP (Amphetamines)– 500ng/ml
 - BUP (Buprenorphine) – 5ng/ml
 - BZO (Benzodiazepines) – 300ng/ml
 - COC (Cocaine) – 150ng/ml
 - mAMP (Methamphetamine)– 1000ng/ml
 - *OXY (Oxycodone)– 100ng/ml
 - OPI (Opiates)– 300ng/ml
 - THC (Cannabis) – 15ng/ml

B. ITEM 2 – 8-PANEL ORAL FLUID TESTERS

1. The oral fluid testers will meet the following specifications:
 - a. The test strips must be single drug per test.
 - b. The oral fluid tester must be an all-inclusive tester and only require the amount of oral fluid needed for the collection and accurate results. No other process or device can be required to perform the test.
 - c. Oral Fluid Test containers must have a saturation indicator strip in the swab of each testing container.
 - d. The oral fluid tester will have easy readable window of test strips and results.

*Change OXY to Oxycodone in II, A, 1, i.